

# Certificate of Conformity

These products were manufactured in accordance with compliance for ISO 13485:2016 and ISO 9001:2015.

## Product Identification

**Name:** CE Grade Water (H<sub>2</sub>O)

**Kit/Reorder number:** C48034

**Storage:** Store at 2°C to 30°C

## Lot Specific Information

Lot Number: M004267

Manufacture Date: 2019-11-11

Expiration Date: 2022-05-11

## Quality Control Results:

Component	Lot Number	Manufacturing Specification	Pass/Fail
CE Grade Water (H <sub>2</sub> O)	M910119	Total Organic Carbon (TOC): ≤ 8 ppm	PASS
		Particulates limit: ≤ 0.22 μm	PASS
		Resistivity: ≥ 16 Megohm	PASS

## Quality Assurance Statement

These reagents were verified and released by SCIEX Quality Control Laboratories and the results above were obtained.

Quality Control Associate: STEPHANIE YAN

Date: June 17, 2020

Signature: \_\_\_\_\_ 

## Support

Technical Support call 1 800-952-4716 or 877-740-2129 or E-mail: [Support@sciex.com](mailto:Support@sciex.com)

This certificate may be obtained at: <http://sciex.com>