

Certificate of Conformity

These products were manufactured in accordance with compliance for ISO 13485:2016 and ISO 9001:2015.

Product Identification

Name: CE Grade Water (H₂O)

Kit/Reorder number: C48034

Storage: Store at 2°C to 30°C

Lot Specific Information

Lot Number:

Manufacture Date:

Expiration Date:

Quality Control Results:

Component	Lot Number	Manufacturing Specification	Pass/Fail
CE Grade Water (H ₂ O)	M912107	Total Organic Carbon (TOC): ≤ 8 ppb	PASS
		Particulates limit: ≤ 0.22 μm	PASS
		Resistivity: ≥ 16 Megohm	PASS


Quality Assurance Statement

These reagents were verified and released by SCIEX Quality Control Laboratories and the results above were obtained.

Quality Control Associate:

Date:

Signature: _____



Support

Technical Support call 1 800-952-4716 or 877-740-2129 or E-mail: Support@sciex.com

This certificate may be obtained at: <http://sciex.com>