



ExionLC™ AD Multiplate Sampler

DECLARATION OF CONFORMANCE

5047299

↑ _
↓ 、

↓ _



DECLARATION OF CONFORMITY
(According to EN ISO/IEC 17050-1:2004)



We

AB Sciex LP

of

71 Four Valley Drive, Concord, ON, Canada L4K 4V8

declare under sole responsibility that the product as originally delivered

Equipment: **LC Auto Sampler**
Product Name: **ExionLC™ AD Multiplate Sampler**

complies with the essential requirements of the following applicable European Directives, and carries the CE marking accordingly:

- 2014/35/EU - Low Voltage Directive**
- 2014/30/EU - Electromagnetic Compatibility Directive**
- 2012/19/EU - Waste Electrical and Electronic Equipment Directive**
- 2011/65/EU - Restriction of Hazardous Substances Directive**

and conforms with the following product standards:

| Category | Standard | Classification |
|-----------------------|---|--|
| Safety | IEC 61010-1:2010 / EN 61010-1:2010 | Equipment Class I, Pollution Degree 2 |
| | IEC 61010-2-081: 2001 +A1:2003 / EN 61010-2-081: 2002 + A1:2003 | |
| | Canada: CAN/CSA-C22.2 No. 61010-1-2012 | |
| | CAN/CSA-C22.2 No. 61010-2-081-04 (R2014) | |
| USA: | ANSI/UL 61010-1:2012 | |
| EMC | IEC 61326-1:2012 / EN 61326-1:2013 | Limit Class A Group 1 ISM 4 kV/8 kV contact/air 3 V/m, 80MHz-1GHz/1.4-2GHz, 1V/m 2-2.7 GHz 1 kV signal lines, 2 kV AC lines 2 kV common mode, 1 kV differential mode 3 V, 0.15-80 MHz 30 A/m, 50/60Hz Dips: 0% for 0.5, 1 cycle; 70% for 25 cycles Short Interruption: 0% for 250 cycles. Class A Group 1 ISM |
| | Reference Standards | |
| | CISPR 11 / EN 55011 | |
| | IEC 61000-4-2 / EN 61000-4-2 | |
| | IEC 61000-4-3 / EN 61000-4-3 | |
| | IEC 61000-4-4 / EN 61000-4-4 | |
| | IEC 61000-4-5 / EN 61000-4-5 | |
| | IEC 61000-4-6 / EN 61000-4-6 | |
| | IEC 61000-4-8 / EN 61000-4-8 | |
| | IEC 61000-4-11 / EN 61000-4-11 | |
| Canada: CISPR 11:2004 | | |

The product was tested in a typical configuration.

Year Mark First Applied: **2015**

We, the undersigned, hereby declare that the equipment specified above conforms to the above Directives and Standards.

Chris Lock
Vice President, LCMS,
Research and Development

Penelope Ernould
Senior Manager,
Quality Management Systems & CDT&R

3rd Oct 2018

Date