



ExionLC™ Degasser

**DECLARATION OF CONFORMANCE**

5047293

↑ \_  
↓ \_

↓ \_



**DECLARATION OF CONFORMITY**  
(According to EN ISO/IEC 17050-1:2004)



We

**AB Sciex LP**

of

**71 Four Valley Drive, Concord, ON, Canada L4K 4V8**

declare under sole responsibility that the product as originally delivered

Equipment: **LC Degasser**  
Product Name: **ExionLC™ Degasser**

complies with the essential requirements of the following applicable European Directives, and carries the CE marking accordingly:

- 2014/35/EU - Low Voltage Directive
- 2014/30/EU - Electromagnetic Compatibility Directive
- 2012/19/EU - Waste Electrical and Electronic Equipment Directive
- 2011/65/EU - Restriction of Hazardous Substances Directive

and conforms with the following product standards:

Category	Standard	Classification
<b>Safety</b>	IEC 61010-1:2010 / EN 61010-1:2010 Australia AS/NZS 61010.1:2003 Canada: CAN/CSA-C22.2 No. 61010-1:2012 New Zealand: AS/NZS 61010.1:2003 USA: ANSI/UL 61010-1:2012	Equipment Class I, Pollution Degree 2
<b>EMC</b>	IEC 61326-1:2013 / EN 61326-1:2013 <b>Reference Standards</b> CISPR 11 / EN 55011 IEC 61000-4-2 / EN 61000-4-2 IEC 61000-4-3 / EN 61000-4-3 IEC 61000-4-4 / EN 61000-4-4 IEC 61000-4-6 / EN 61000-4-6 IEC 61000-4-8 / EN 61000-4-8 IEC 61000-4-11 / EN 61000-4-11  Australia/New Zealand: AS/NZS CISPR 11:2011 Canada: CISPR 11:2004	<b>Limit</b> Class A Group 1 ISM 4 kV/8 kV contact/air 3 V/m, 80MHz-1GHz/1.4-2GHz, 1V/m 2-2.7 GHz 0.5 kV signal lines 3 V, 0.15-80 MHz 30 A/m, 50/60Hz Dips: 0% for 0.5, 1 cycle; 70% for 25 cycles Short Interruption: 0% for 250 cycles. Class A Group 1 ISM Class A Group 1 ISM

The product was tested in a typical configuration.

Year Mark First Applied: **2015**

We, the undersigned, hereby declare that the equipment specified above conforms to the above Directives and Standards.

Chris Lock  
Vice President, LCMS,  
Research and Development

Penelope Ernould  
Senior Manager,  
Quality Management Systems & CDT&R

3<sup>rd</sup> Oct 2018

Date