



ExionLC™ UV Detector

**DECLARATION OF CONFORMANCE**

5042658

↑ \_  
↓ \_

↓ \_



**DECLARATION OF CONFORMITY**  
(According to EN ISO/IEC 17050-1:2004)



We

**AB Sciex LP**

of

**71 Four Valley Drive, Concord, ON, Canada L4K 4V8**

declare under sole responsibility that the product as originally delivered

Equipment: **UV – VIS Spectrophotometric Detector**  
Product Name: **ExionLC™ UV Detector**

complies with the essential requirements of the following applicable European Directives, and carries the CE marking accordingly:

- 2014/35/EU - Low Voltage Directive**
- 2014/30/EU - Electromagnetic Compatibility Directive**
- 2012/19/EU - Waste Electrical and Electronic Equipment Directive**
- 2011/65/EU - Restriction of Hazardous Substances Directive**


and conforms with the following product standards:

<b>Category</b>	<b>Standard</b>	<b>Classification</b>
<b>Safety</b>	IEC 61010-1:2010 / EN 61010-1:2010	Equipment Class I, Pollution Degree 2
	Australia: AS 61010.1:2003	
	Canada: CAN/CSA-C22.2 No. 61010-1:2012	
	New Zealand: AS 61010.1:2003	
	USA: ANSI/UL 61010-1:2012	
<b>EMC</b>	IEC 61326-1:2012 / EN 61326-1:2013	<b>Limit</b> Class A Group 1 ISM 4 kV/8 kV contact/air 3 V/m, 80MHz-1GHz/1.4-2GHz, 1V/m 2-2.7 GHz 1 kV signal lines, 2 kV AC lines 2 kV common mode, 1 kV differential mode 3 V, 0.15-80 MHz 30 A/m, 50/60Hz Dips: 0% for 0.5, 1 cycle; 70% for 25 cycles Short Interruption: 0% for 250 cycles. Class A Group 1 ISM Class A Group 1 ISM
	<b>Reference Standards</b>	
	CISPR 11 / EN 55011	
	IEC 61000-4-2 / EN 61000-4-2	
	IEC 61000-4-3 / EN 61000-4-3	
	IEC 61000-4-4 / EN 61000-4-4	
	IEC 61000-4-5 / EN 61000-4-5	
	IEC 61000-4-6 / EN 61000-4-6	
	IEC 61000-4-8 / EN 61000-4-8	
	IEC 61000-4-11 / EN 61000-4-11	
	Australia/New Zealand: AS/NZS CISPR 11:2011	
	Canada: CISPR 11:2004	


The product was tested in a typical configuration.

Year Mark First Applied: **2016**

We, the undersigned, hereby declare that the equipment specified above conforms to the above Directives and Standards.

  
Gordon Logan

Senior Vice President, Research and Development

  
Michael McDonald

Senior Manager, Quality Assurance

2016/01/14  
Date