



SCIEX Triple Quad™ 7500 LC-MS/MS System – QTRAP® Ready MASS SPECTROMETER
(SCIEX Triple Quad™ 7500 QTR)

DECLARATION OF CONFORMITY

5080782

↑ _

↓ 、

↓ _



DECLARATION OF CONFORMITY
(According to EN ISO/IEC 17050-1:2017)



We **AB Sciex LP**
of **71 Four Valley Drive, Concord, ON, Canada L4K 4V8**

declare under sole responsibility that the product as originally delivered

Equipment: **Liquid Chromatography-Mass Spectrometer/Mass Spectrometer**
Product Name: **SCIEX Triple Quad™ 7500 LC-MS/MS System – QTRAP® Ready MASS SPECTROMETER (SCIEX Triple Quad™ 7500 QTR)**
Equipment Components: **Mass Spectrometer, Ion Source**
Product Options: **OptiFlow® Pro Ion Source with TurbolonSpray™ (TIS) or Atmospheric Pressure Chemical Ionization (APCI)**

complies with the essential requirements of the following applicable European Directives, and carries the CE marking accordingly:

- 2014/35/EU - Low Voltage Directive**
- 2014/30/EU - Electromagnetic Compatibility Directive**
- 2012/19/EU - Waste Electrical and Electronic Equipment Directive**
- 2011/65/EU - Restriction of Hazardous Substances Directive**

and conforms with the following product standards:

Category	Standard	Classification
Safety	EU: IEC 61010-1:2010 / EN 61010-1:2010 + Corrigendum 1:2011+ Amd 1: 2016 IEC 61010-2-061:2015 / EN 61010-2-061:2015	Equipment Class I, Pollution Degree 2
	Australia AS 61010.1:2003, IEC 61010-2-061:2005	
	New Zealand: AS 61010.1:2003, IEC 61010-2-061:2005	
	Canada: CAN/CSA C22.2 No. 61010-1-12 + Amd 1: 2018 CAN/CSA C22.2 No. 61010-2-061-19	
	USA: UL 61010-1 (Ed. 3) + Amd 1: 2018, UL 61010-2-061 (Ed. 4)	
EMC	EU: IEC 61326-1:2012 / EN 61326-1:2013 Reference Standards: CISPR 11 / EN 55011 IEC 61000-4-2 / EN 61000-4-2 IEC 61000-4-3 / EN 61000-4-3 IEC 61000-4-4 / EN 61000-4-4 IEC 61000-4-5 / EN 61000-4-5 IEC 61000-4-6 / EN 61000-4-6 IEC 61000-4-8 / EN 61000-4-8 IEC 61000-4-11 / EN 61000-4-11	Limit Class A Group 1 ISM 4 kV/8 kV contact/air 3 V/m, 80MHz-1GHz/1.4-2GHz, 1V/m 2-2.7 GHz 0.5 kV signal lines, 1 kV AC lines 0.5 kV differential mode, 1 kV common mode 3 V, 0.15-80 MHz 30 A/m, 50/60Hz Dips: 0% for 0.5, 1 cycle; 70% for 25 cycles Short Interruption: 0% for 250 cycles. Class A Group 1 ISM
	Australia/New Zealand: AS/NZS CISPR 11:2011	
	Canada: ICES-001 (CISPR 11:2004)	
	USA: FCC Part 15	
		Class A Group 1 ISM Class A

The product was tested in a typical configuration.

Year Mark First Applied: **2020** Replaced: **N/A**

We, the undersigned, hereby declare that the equipment specified above conforms to the above Directives and Standards.

5/12/2020

5/12/2020

X

X Nichole Riek

Signed by: Lock, Chris
Chris Lock
Vice President, LCMS
Research and Development

Signed by: Riek, Nichole
Nichole Riek
Vice President,
Quality Assurance and
Regulatory Affairs



AB Sciex Netherlands B.V.
1e Tochtweg 11,
2913LN Nieuwerkerk aan den IJssel
Netherlands