



SCIEX X500R QTOF and SCIEX X500B QTOF
MASS SPECTROMETER

DECLARATION OF CONFORMANCE

5057665





DECLARATION OF CONFORMITY
(According to EN ISO/IEC 17050-1:2004)



We
of
AB Sciex LP
71 Four Valley Drive, Concord, ON, Canada L4K 4V8

declare under sole responsibility that the product as originally delivered

Equipment: **Mass Spectrometer**
Product Name: **SCIEX X500R QTOF and SCIEX X500B QTOF**
Product Options:
Turbo V™ or IonDrive™ Turbo V Ion Source with TurbolonSpray® (TIS) or Atmospheric Pressure Chemical Ionization (APCI)

complies with the essential requirements of the following applicable European Directives, and carries the CE marking accordingly:

- 2014/35/EU - Low Voltage Directive
- 2014/30/EU - Electromagnetic Compatibility Directive
- 2012/19/EU - Waste Electrical and Electronic Equipment Directive
- 2011/65/EU - Restriction of Hazardous Substances Directive

and conforms with the following product standards:

Category	Standard	Classification
Safety	IEC 61010-1:2010 / EN 61010-1:2010	Equipment Class I, Pollution Degree 2
	IEC 61010-2-061:2015	
	IEC 61010-1:2001 / EN 61010-1:2001	
	IEC 61010-2-061:2005 / EN 61010-2-061:2005	
	Australia: AS 61010.1:2003, IEC 61010-2-061:2005	
	Canada: CAN/CSA C22.2 No. 61010-1-12	
	CAN/CSA C22.2 No. 61010-2-061-15	
	New Zealand: AS 61010.1:2003, IEC 61010-2-061:2005	
	USA: ANSI/UL 61010-1-2012, IEC 61010-2-061-15	
	EMC	
Reference Standards		
CISPR 11 / EN 55011		
IEC 61000-4-2 / EN 61000-4-2		
IEC 61000-4-3 / EN 61000-4-3		
IEC 61000-4-4 / EN 61000-4-4		
IEC 61000-4-5 / EN 61000-4-5		
IEC 61000-4-6 / EN 61000-4-6		
IEC 61000-4-8 / EN 61000-4-8		
IEC 61000-4-11 / EN 61000-4-11		
IEC 61000-3-2:2005+A1:2008+A2:2009 / EN 61000-2-3:2006+A1:2009+A2:2009		
IEC 61000-3-3:2013 / EN61000-3-3:2013		
Australia/New Zealand: AS/NZS CISPR 11:2011	Class A Group 1 ISM	
Canada: CISPR 11:2004		

The product was tested in a typical configuration.

Year Mark First Applied: **2015**

We, the undersigned, hereby declare that the equipment specified above conforms to the above Directives and Standards.

Gordon Logan

Senior Vice President, Research and Development

Penelope Ernould

Senior Manager,
Quality Management Systems and CDT&R

Date