



API 4000™ LC/MS/MS System,
4000 QTRAP® LC/MS/MS System
and AB Sciex API 4000+ MASS SPECTROMETER

DECLARATION OF CONFORMANCE

5047963





DECLARATION OF CONFORMITY
(According to EN ISO/IEC 17050-1:2004)



We
of
AB Sciex LP
71 Four Valley Drive, Concord, ON, Canada L4K 4V8

declare under sole responsibility that the product as originally delivered

Equipment: **Liquid Chromatography/Mass Spectrometer/Mass Spectrometer**
Product Name: **API 4000™ LC/MS/MS System, 4000 QTRAP® LC/MS/MS System and AB Sciex API 4000+**
Product Options: **Turbo V™ Ion Source with TurbolonSpray™ (TIS) or Atmospheric Pressure Chemical Ionization (APCI) probes; DuoSpray™ and NanoSpray® III Ion Sources.**

complies with the essential requirements of the following applicable European Directives, and carries the CE marking accordingly:

- 2014/35/EU - Low Voltage Directive
- 2014/30/EU - Electromagnetic Compatibility Directive
- 2012/19/EU - Waste Electrical and Electronic Equipment Directive

and conforms with the following product standards:

Category	Standard	Classification	
Safety	IEC 61010-1:2001 / EN 61010-1:2001	Equipment Class I, Pollution Degree 2	
	IEC 61010-2-061:2003 / EN 61010-2-061:2003		
	Australia: AS 61010.1:2003, IEC 61010-2-061:2005		
	Canada: CAN/CSA C22.2 No. 61010-1-04 CAN/CSA C22.2 No. 61010-2-061-04		
	New Zealand: AS 61010.1:2003, IEC 61010-2-061:2005 USA: ANSI/UL 61010-1:2004, IEC 61010-2-061-2003		
EMC	IEC 61326-1:2012 / EN 61326-1:2013	Limit Class A Group 1 ISM 4 kV/8 kV contact/air 3 V/m, 80MHz-1GHz/1.4-2GHz, 1V/m 2-2.7 GHz 0.5 kV signal lines, 1 kV AC lines 0.5 kV differential mode, 1 kV common mode 3 V, 0.15-80 MHz 30 A/m, 50/60Hz Dips: 0% for 0.5, 1 cycle; 70% for 25 cycles Short Interruption: 0% for 250 cycles.	
	Reference Standards		
	CISPR 11 / EN 55011		
	IEC 61000-4-2 / EN 61000-4-2		
	IEC 61000-4-3 / EN 61000-4-3		
	IEC 61000-4-4 / EN 61000-4-4		
	IEC 61000-4-5 / EN 61000-4-5		
	IEC 61000-4-6 / EN 61000-4-6		
	IEC 61000-4-8 / EN 61000-4-8		
	IEC 61000-4-11 / EN 61000-4-11		
	Australia/New Zealand: AS/NZS CISPR 11:2011		
	Canada: ICES-001 (CISPR 11:2004)		
	USA: FCC Part 15		
			Class A Group 1 ISM
			Class A Group 1 ISM
	Class A		

The product was tested in a typical configuration.

Year Mark First Applied: **1996**

We, the undersigned, hereby declare that the equipment specified above conforms to the above Directives and Standards.

Gordon Logan

Senior Vice President, Research and Development

Elizabeth Culotta

Director,
Global Product Quality Management

Date