



SCIEX SelexION® Series ION MOBILITY SYSTEM

DECLARATION OF CONFORMANCE

5058359

↑
↓

↓



DECLARATION OF CONFORMITY
(According to EN ISO/IEC 17050-1:2004)



We
of
AB Sciex LP
71 Four Valley Drive, Concord, ON, Canada L4K 4V8

declare under sole responsibility that the product as originally delivered

Equipment: **Ion Mobility System**
Product Name: **SelexION®**
Product Options: For use with
LC/MS/MS Mass Spectrometer, Models SCIEX QTRAP® 5500, SCIEX Triple Quad™ 5500,
SCIEX QTRAP® 6500, SCIEX Triple Quad™ 6500, Triple TOF™ 5600+ and Triple TOF™ 6600
Product Name: **SelexION® +**
Product Options: For use with LC/MS/MS Mass Spectrometer, Models SCIEX QTRAP® 6500+, SCIEX Triple
Quad™ 6500+

complies with the essential requirements of the following applicable European Directives, and carries the CE marking accordingly:

- 2014/35/EU - Low Voltage Directive
- 2014/30/EU - Electromagnetic Compatibility Directive
- 2012/19/EU - Waste Electrical and Electronic Equipment Directive
- 2011/65/EU - Restriction of Hazardous Substances Directive

and conforms with the following product standards:

| | | |
|----------------------------------|---|---|
| Category Safety | Standard IEC 61010-1:2010 / EN 61010-1:2010 Canada: CAN/CSA C22.2 No. 61010-1-12 USA: ANSI/UL 61010-1:2012 | Classification Equipment Class I, Pollution Degree 2 |
| EMC | IEC 61326-1:2012 / EN 61326-1:2013 Reference Standards CISPR 11 / EN 55011 IEC 61000-4-2 / EN 61000-4-2 IEC 61000-4-3 / EN 61000-4-3 IEC 61000-4-4 / EN 61000-4-4 IEC 61000-4-5 / EN 61000-4-5 IEC 61000-4-6 / EN 61000-4-6 IEC 61000-4-8 / EN 61000-4-8 IEC 61000-4-11 / EN 61000-4-11 Canada: ICES-001 (CISPR 11:2004) USA: FCC Part 15 | Limit Class A Group 1 ISM 4 kV/8 kV contact/air 3 V/m, 80MHz-1GHz/1.4-2GHz, 1V/m 2-2.7 GHz 0.5 kV signal lines, 1 kV AC lines 0.5 kV differential mode, 1 kV common mode 3 V, 0.15-80 MHz 30 A/m, 50/60Hz Dips: 0% for 0.5, 1 cycle; 70% for 25 cycles Short Interruption: 0% for 250 cycles. Class A Group 1 ISM Class A |

The product was tested in a typical configuration.

Year Mark First Applied: **2011**

We, the undersigned, hereby declare that the equipment specified above conforms to the above Directives and Standards.

Gordon Logan

Senior Vice President, Research and
Development

Penelope Arnould

Senior Manager,
Quality Management Systems and CDT&R

2017/06/30
Date