



SCIEX Triple TOF™ 5600+ MASS SPECTROMETER

DECLARATION OF CONFORMANCE

5062671

↑
↓

↓



DECLARATION OF CONFORMITY
(According to EN ISO/IEC 17050-1:2004)



We

AB Sciex LP

of

71 Four Valley Drive, Concord, ON, Canada L4K 4V8

declare under sole responsibility that the product as originally delivered

Equipment: **Mass Spectrometer**

Product Name: **SCIEX Triple TOF™ 5600+**

Product Options: DuoSpray™ with TurbolonSpray™ (TIS) or Atmospheric Pressure Chemical Ionization (APCI); Turbo V™ with TurbolonSpray™ (TIS) or Atmospheric Pressure Chemical Ionization (APCI); NanoSpray® III; or Digital PicoView® Nanospray DPV-450 Ion Sources; SelexION® Ion Mobility System

complies with the essential requirements of the following applicable European Directives, and carries the CE marking accordingly:

- 2014/35/EU - Low Voltage Directive
- 2014/30/EU - Electromagnetic Compatibility Directive
- 2012/19/EU - Waste Electrical and Electronic Equipment Directive
- 2011/65/EU - Restriction of Hazardous Substances Directive

and conforms with the following product standards:


Category	Standard	Classification
Safety	IEC 61010-1:2010 / EN 61010-1:2010 + Corrigendum 1:2011	Equipment Class I, Pollution Degree 2
	IEC 61010-2-061:2015 / EN 61010-2-061:2015	
	Australia: AS 61010.1:2003, IEC 61010-2-061:2005	
	Canada: CAN/CSA C22.2 No. 61010-1-12 CAN/CSA C22.2 No. 61010-2-061-15	
	New Zealand: AS 61010.1:2003, IEC 61010-2-061:2005	
USA: UL 61010-1 (Ed. 3), UL 61010-2-061 (Ed. 3)		
EMC	IEC 61326-1:2012 / EN 61326-1:2013	Limit Class A Group 1 ISM 4 kV/8 kV contact/air 3 V/m, 80MHz-1GHz/1.4-2GHz, 1V/m 2-2.7 GHz 0.5 kV signal lines, 1 kV AC lines 0.5 kV differential mode, 1 kV common mode 3 V, 0.15-80 MHz 30 A/m, 50/60Hz Dips: 0% for 0.5, 1 cycle; 70% for 25 cycles Short Interruption: 0% for 250 cycles.
	Reference Standards	
	CISPR 11 / EN 55011	
	IEC 61000-4-2 / EN 61000-4-2	
	IEC 61000-4-3 / EN 61000-4-3	
	IEC 61000-4-4 / EN 61000-4-4	
	IEC 61000-4-5 / EN 61000-4-5	
	IEC 61000-4-6 / EN 61000-4-6	
	IEC 61000-4-8 / EN 61000-4-8	
	IEC 61000-4-11 / EN 61000-4-11	
Australia/New Zealand: AS/NZS CISPR 11:2011	Class A Group 1 ISM	
Canada: ICES-001 (CISPR 11:2004)		Class A Group 1 ISM
USA: FCC Part 15		Class A


The product was tested in a typical configuration.

Year Mark First Applied: **2012**

Replaces: **5061234**

We, the undersigned, hereby declare that the equipment specified above conforms to the above Directives and Standards.


Chris Lock
Vice President, LCMS
Research and Development


Penelope Ernould
Senior Manager,
Quality Management Systems and CDT&R

2018-03-12
Date