

Compliant attribute monitoring for biopharmaceutical product quality attributes employing intact mass analysis

Application of intact mass attribute monitoring workflow in SCIEX OS Software 1.7

Zoe Zhang, Sean McCarthy
SCIEX, USA,

Liquid chromatography mass spectrometry (LC-MS) is a widely adopted analytical technique used in biopharmaceutical research, as it enables detection and measurement of a variety of product quality attributes (PQAs). Intact mass and subunit analysis are among of the most common LC-MS assays employed in research and development due to the high degree of information provided and high sample throughput. They have been used for determination of molecular weight, glycoform distribution, other post translational modification (PTM) assessment, and domain specific information.

In recent years, multiple attribute methodology (MAM) has gained popularity in monitoring biologic PQAs of biotherapeutics within late-stage development and manufacturing laboratories.¹ The concept of MAM is to offer a single method capable of monitoring various PQAs simultaneously. This approach may increase efficiency and quality by replacing or supplementing conventional QC tests. Current MAM solutions require the generation of peptides from a protein digest, which increases the potential of inducing PTMs during sample manipulation and is time consuming. With the advantage of limited sample preparation and the potential for real time monitoring, the implementation of MAM for intact mass analysis (including subunit analysis) has attracted high interest in biopharmaceutical industry.



Essential to successful implementation of a thorough and productive intact MAM is a software solution that can manage all aspects of the workflow, including: product quality attribute (PQA) definition, tracking, and quantification in a compliant environment. To date, there has not been a user friendly or streamlined software package capable of data acquisition, intact mass reconstruction and intact mass MAM analysis for compliant environments. With the launch of SCIEX OS Software 1.7 from SCIEX, an integrated software solution from data acquisition, data analysis to intact mass MAM is now available. In this technical note, the use of SCIEX OS Software 1.7 to perform an intact MAM analysis is demonstrated, highlighting a high throughput analysis of intact data as well as the capability to monitor and track quality attributes from different lots in a compliant environment.




Figure 1. SCIEX OS Software results table. Results table showing extracted ion chromatogram (XIC), TOF-MS raw spectra, reconstructed spectra, percentage calculation and statistics graph in SCIEX OS Software 1.7.

Key features of the MAM Solution in SCIEX OS Software

- Complete and compliant software solution for intact MAM workflows with high resolution mass spectrometry
- Powerful product attribute definition, tracking, and quantitation with flexible custom calculations for attribute-level assessment based on specific user needs
- Reconstruction level quantification for high specificity analysis and comparison between raw and reconstructed data for rapid investigation of similarities and differences in samples

Methods

Sample preparation: Adalimumab antibody samples were diluted to a concentration of 0.1 mg/mL using deionized water for intact mass analysis.

Chromatography: Separation was accomplished using an ExionLC™ System fitted with an Agilent PLRP-S column (2.1 mm×50 mm, 300 Å, 5 µm) at 80°C using the gradient shown in Table 1. Mobile phase A was 0.1% formic acid in water and mobile phase B was 0.1% formic acid in acetonitrile.

Table 1. LC conditions for intact mass analysis.

Time [min]	%A	%B	Flow Rate [ml/min]
Initial	85	15	0.5
5.0	85	15	0.5
9.0	5	95	0.5
11.4	5	95	0.5
11.5	85	15	0.5
15.0	85	15	0.5

Mass spectrometry: A SCIEX X500B QTOF System with a Turbo V™ Ion Source fitted with a Twin Sprayer ESI Probe was used for data acquisition. MS instrument conditions are listed in Table 2.

Data Processing: Data were processed using the Analytics tool in SCIEX OS Software 1.7.




Figure 2. Reconstruction of intact mass analysis in SCIEX OS Software 1.7. (Top) Total ion chromatogram shows a single major peak, the TOF-MS raw spectra (middle) is extracted from the apex of that peak. Reconstruction of the TOF-MS raw spectra to a zero charge state (bottom).

Table 2. MS Parameters.

Parameter	Setting
Scan Mode	Positive
Gas 1	60 psi
Gas 2	60 psi
Curtain Gas	30 psi
Temperature	550°C
Ion Spray Voltage	5500 V
Time Bins to Sum	80
Accumulation Time	0.5 sec
TOF Start Mass	m/z 400
TOF Stop Mass	m/z 4000
Declustering Potential	250 V
Collision Energy	10

Attribute Definition

Data was acquired using SCIEX OS Software with a SCIEX X500B QTOF System. Acquired data was processed using the built in reconstruction algorithm without the need for definition of the target sequence (Figure 2). For studies in which target sequences may be refined or where the scientific question relates only to PTMs which are easily identified using mass shifts, such as glycosylation, this ability streamlines data analysis.

From the data, targeted attributes are easily defined within SCIEX OS Software 1.7 as shown in Figure 3. The expected molecular weight, m/z range, retention time (RT) and reconstruction range are entered based on the study. If the some relevant information is not known when developing the assay, such as RT and expected molecular weight, these may also be defined during integration by selecting peaks from chromatographic and reconstructed spectra.

Integration parameters for individual components can be adjusted to achieve accurate quantitation in SCIEX OS Software 1.7, providing the flexibility to refine processing parameters on a component by component level (Figure 4). Using the advanced reconstruction parameters, the user can define the number of iterations, S/N threshold, step mass size, and mass ranges for the assay.

Within each attribute group, the attribute level can be calculated using flexible custom calculations. Figure 5 and Figure 6 show examples of percentage calculation for glycoform distribution.

Figure 5 represents the equation defined to sum the peak areas of all targeted glycoforms, while Figure 6 shows the percentage calculation based on the summed data. After processing, the calculated values are displayed in a results table as shown in Figure 1. Within minutes the results are generated and available for review. There is no practical limit to the number of attributes that can be defined or monitored as part of an assay, highlighting that SCIEX OS Software 1.7 has the flexibility to monitor a large number of attributes simultaneously.

[MQ4] Modify Method

Workflow Select or verify the analyte and internal standard names and masses.

Components Import Export... Options

Row	IS	Group	Name	Expected MW (Da)	m/z Range for XIC (Da)	Retention Time (min)	Reconstruction Start Mass (Da)	Reconstruction Stop Mass (Da)	IS Name	Experiment Index
1	<input type="checkbox"/>	2Lysloss	G0F/Man5	147857.18	400 - 4000	6.54	145000.00	155000.00		1 +TOF MS (400 - 4000)
2	<input type="checkbox"/>	2Lysloss	G0F-GlcNAc/G0F	147877.00	400 - 4000	6.54	145000.00	155000.00		1 +TOF MS (400 - 4000)
3	<input type="checkbox"/>	2Lysloss	G0F-GlcNAc/G1F	148031.54	400 - 4000	6.54	145000.00	155000.00		1 +TOF MS (400 - 4000)
4	<input type="checkbox"/>	2Lysloss	G0F/G0F	148082.15	400 - 4000	6.54	145000.00	155000.00		1 +TOF MS (400 - 4000)
5	<input type="checkbox"/>	2Lysloss	G0F/G1F	148244.45	400 - 4000	6.54	145000.00	155000.00		1 +TOF MS (400 - 4000)
6	<input type="checkbox"/>	2Lysloss	G1F/G1F	148405.51	400 - 4000	6.54	145000.00	155000.00		1 +TOF MS (400 - 4000)
7	<input type="checkbox"/>	2Lysloss	G1F/G2F	148569.27	400 - 4000	6.54	145000.00	155000.00		1 +TOF MS (400 - 4000)
8	<input type="checkbox"/>									

Figure 3. Attribute definition for intact mass analysis in SCIEX OS Software.

[MQ4] Modify Method

Workflow For each component, configure the parameters to optimize peak integration

Components Options

Integration Algorithm: MQ4

Library Search

Calculated Columns

Flagging Rules

Advanced

Formula Finder

Non-targeted Peaks

Retention Time (RT)

Expected RT: 6.54 min

XIC Gaussian Smooth Width: 0.0 points

Recentered on the largest XIC Peak

RT Half Window: 30.0 sec

Number of Spectra to Average: 21

Reconstruction

Resolution: 5000

Reconstruction Start Mass: 145000.00 Da

Reconstruction Stop Mass: 155000.00 Da

Advanced Reconstruction

Number of Iterations: 20

Signal to Noise threshold: 10

Step Mass: 1.00 Da

Use input m/z Range

Start m/z: Da

Stop m/z: Da

Apply RT, Reconstruction and Adv.Reconstruction parameters to: Component

Mass Peak Selection

Expected MW: 148082.15 Da

Mass Half Window: 5 Da

Apply

G0F/G0F, XIC from 20191015 Adalimumab lot 3.wiff2 (sa...1) - 20191015 Adalimumab lot 3, +TOF MS (400 - 4000) m/z: 400.0000 - 4000.0000 Da, RT: 6.57 min

Spectrum from 20191015 Adalimumab lot 3.wiff2 (sample...b lot 3, +TOF MS (400 - 4000) from 6.496 to 6.638 min

G0F/G0F, Reconstruction from 20191015 Adalimumab lot ... lot 3, +TOF MS (400 - 4000) from 6.496 to 6.638 min MW: 148081.87 Da, Area: 6.540e5, Height: 29168.684

Process & Close Close Help

Figure 4. Integration parameter definitions in SCIEX OS Software.

Use the calculator to create a new formula.

Formula name

$= \text{GETGROUP}([\text{Area}];1)+\text{GETGROUP}([\text{Area}];2)+\text{GETGROUP}([\text{Area}];3)+$

COUNT	MAX	STDEV	Clear
SUM	MIN	MEDIAN	(
MEAN	ABS	MAD)
/	*	-	+

▼ Regression parameters
 r
 r²
 Slope
 Intercept
 Quadratic coefficient
 Linear coefficient
 Constant term
 ▼ Columns

Figure 5. Custom calculations within attribute group. Calculation formula showing the summing peak areas from different attributes within the same attribute group.

Use the calculator to create a new formula.

Formula name

$= [\text{Area}]/[\text{SumArea}]*100$

COUNT	MAX	STDEV	Clear
SUM	MIN	MEDIAN	(
MEAN	ABS	MAD)
/	*	-	+

▼ Regression parameters
 r
 r²
 Slope
 Intercept
 Quadratic coefficient
 Linear coefficient
 Constant term
 ▼ Columns

Figure 6. Custom calculations from summed data. Calculation of percentage in SCIEX OS Software.

For each attribute defined in an assay, the acceptance criteria for that attribute can also be defined. Each attribute acceptance level is defined independently providing flexibility for specific assays. To highlight outliers, flagging criteria may be set based on a percentage variation from a defined value or over a range of values. The pass/fail range for each attribute can be aggregated and used to determine the overall pass/fail status for a corresponding attribute, allowing users to screen values that don't meet the necessary criteria. For example, in the presented assay the acceptance range of GOF was set to a range of 37-45% with results falling outside of this range being flagged (Figure 7).

Attribute Quantification

Sample batches can be submitted for processing in SCIEX OS Software. During processing, the targeted attributes are quantified and tracked across each submitted sample. Sample differences related to the defined attributes are easily identified. By design, SCIEX OS Software can process batches of two or greater files simultaneously to enable direct sample comparison. For each sample and study, a summary table is generated (Figure 1). The summary table provides an overview of the calculated attribute levels and a pass/fail indication. The results may be viewed in totality or for each specific attribute defined among the samples. Failed attributes are highlighted in red, as demonstrated in Figure 7 and the results can be easily filtered to focus on only those responses which do not meet the defined acceptance criteria.

Attributes can be examined directly by viewing the underlying raw and reconstructed data used for each calculation by selected entries in the table generated in SCIEX OS Software, which allows independent data validation within the same software interface. This same table provides direct access to extracted ion chromatograms, raw MS spectra, and reconstructed spectra for the selected component, which enables users to perform attribute monitoring, data investigation, and assay refinement seamlessly within the same workspace. The optimized parameters can be applied to the entire assay within results table interface or to specific attributes or responses. Finally, a metric plot can be generated for the results as a visualization tool, so the review of attribute response across different samples is expedited.

Conclusions

The SCIEX OS Software 1.7 provides a breakthrough in intact MAM analysis by providing a streamlined and compliant software package, from data acquisition through data analysis. The entire intact mass MAM solution is provided including survey analysis, attribute definition, custom calculations, and method optimization. The ability of SCIEX OS Software 1.7 to offer advanced integration for individual components enables accurate quantitation, which is essential for MAM assays. The full compliance feature set offered by SCIEX OS Software provides the potential to implement intact mass MAM in regulated environments. SCIEX OS Software provides a superior MAM solution which can fulfill the needs in discovery, development, and quality control laboratories.




Figure 7. Attribute summary. Attributes are summarized in tables with each entry linking to underlying XIC, raw, and reconstructed data. The trend of each attribute is plotted as shown in the middle panel to expedite data review.

References

1. Rogers RS et al. (2018). A View on the Importance of “Multi-Attribute Method” for Measuring Purity of Biopharmaceuticals and Improving Overall Control Strategy. [The AAPS Journal, 20, Article 7.](#)

The SCIEX clinical diagnostic portfolio is For In Vitro Diagnostic Use. Rx Only. Product(s) not available in all countries. For information on availability, please contact your local sales representative or refer to <https://sciex.com/diagnostics>. All other products are For Research Use Only. Not for use in Diagnostic Procedures.

Trademarks and/or registered trademarks mentioned herein, including associated logos, are the property of AB Sciex Pte. Ltd. or their respective owners in the United States and/or certain other countries.

© 2020 DH Tech. Dev. Pte. Ltd. RUO-MKT-02-11314-A. AB SCIEX™ is being used under license.