

Impact of High-Resolution MRM Analysis on Complex Proteomes

Targeted Protein Quantification using MRM^{HR} Workflow on the TripleTOF[®] Systems

Kathleen Lewis, Christie Hunter
SCIEX, USA

Multiple Reaction Monitoring (MRM) on triple quadrupole based instrumentation is the standard mass spectrometric technique of choice for targeted MS quantification in all application areas, due to its ability to provide the highest specificity and sensitivity for the detection of specific components in complex mixtures. However, the speed and sensitivity of today's accurate mass MS systems have enabled a new quantification strategy with similar performance characteristics. In this strategy (termed MRM^{HR} workflow or parallel reaction monitoring, PRM), looped MS/MS spectra are collected at high-resolution with short accumulation times, and then fragment ions are extracted post-acquisition to generate MRM-like peaks for integration and quantification. With instrumentation like the TripleTOF[®] Systems, this targeted technique is sensitive and fast enough to enable quantitative performance similar to higher end triple quadrupole instruments, with full fragmentation data measured at high resolution and high mass accuracy.

Previous work has shown that MRM analysis has the power to detect a broad dynamic range of proteins down to less than 50 protein copies/cell in a complex, unfractionated sample of a yeast cell lysate digest¹. Here, we used the same model lysate to explore the effect of the high resolution capabilities of the TripleTOF[™] 5600 System in MRM^{HR} mode (Figure 1) on a set of proteins representing the lower end of cellular abundances (<50-10,000 copies/cell). We explored the impact of various data



processing strategies on quantitative quality.

Key Features of the MRM^{HR} Workflow on TripleTOF[®] Systems

- TripleTOF[®] Systems have the MS/MS sensitivity and speed to perform the high resolution MRM-like workflow- MRM^{HR}
- Fast MS/MS acquisition rates ensure sufficient points across the LC peak for good quantitative reproducibility
- Full scan MS/MS enables post-acquisition optimization of quantitative performance by selecting the best fragment ions and the best resolution for generation of extracted ion chromatograms (XICs)





Figure 1. Looped High Resolution MS/MS – MRM^{HR}. (Left) In this analysis, the instrument is set up such that Q1 is fixed on a target peptide m/z, the peptide is fragmented in the collision cell and the full scan TOF MS/MS is acquired. (Right) This is repeated across the LC run. After data acquisition, extracted ion chromatograms on sequence specific ions are generated. Multiple fragment ions can be monitored or even summed together because the full scan MS/MS is always acquired. After this extraction, the processing for quantification is now very similar to how one would handle MRM data acquired on a triple quad or QTRAP[®] system. Multiple target peptides can be monitored per LC run by looped together multiple different MS/MS experiments, and these can even be scheduled in time to allow larger number of peptides to be analyzed.

Materials and Methods

Sample Preparation: The yeast cell lysate was denatured using RapiGest, then reduced with tris-(2-carboxyethyl)phosphine and alkylated with iodoacetamide. An overnight digestion with trypsin was performed (37 °C).

Chromatography: The sample was analyzed using the NanoLC Ultra™ 2D System plus combined with the cHiPLC® system configured in trap elute mode (SCIEX). The cell lysate (2 µg total protein) was loaded at 2 µL/min onto a 200 µm x 0.5 mm precolumn and eluted at 300 nL/min onto an analytical 75 µm x 150 mm column. Both pre-column and analytical column were filled with ChromXP C18-CL 3µm 120 Å phase (SCIEX). An elution gradient of 5-35% acetonitrile (0.1% formic acid) with a 60 min gradient was used. A µL pick-up method was used for the autosampler method.

Mass Spectrometry: Multiplexed looped MS/MS acquisition methods (MRM^{HR} methods) targeting proteins spanning a range of abundances from 10,000-<50 copies/cell were developed using Skyline software. Data was acquired on the TripleTOF® 5600 System, where Q1 resolution was set to Unit and CE was computed for each precursor. High sensitivity MS/MS mode was used, to obtain MS/MS spectra with resolution >15,000 (FWHM). Accumulation times were varied and replicates were run of each experimental condition.

Data Processing: Data was processed using MultiQuant™ Software, where multiple fragment ions were extracted at variable extraction widths.




Figure 2. Abundance Curve of Yeast. Yeast was chosen as a model organism because of the well characterized dynamic range (yellow)². The blue trace represents proteins identified in 60 min LC gradient using a standard protein identification workflow. Because higher abundance proteins are easily quantified by this technique, we chose proteins from <50 to 10,000 copies per cell target with the MRM^{HR} workflow (pink circles).

Dynamic Range

As shown in Figure 2, quantification was attempted on lower abundance proteins in the yeast cell lysate to explore the various data collection and processing options for the MRM^{HR} workflow on the TripleTOF System. As illustrated in Figure 4, very good reproducibility was achieved even on these lower abundance proteins using this strategy.




Figure 3. High Resolution MS/MS of Peptide MRI1.VLDQLLLPYTTK. Looped full scan MS/MS is acquired across the LC peak at high resolution, >15,000 or >30,000, as shown in the inset for the y7 fragment ion. High resolution XICs are generated post-acquisition and are typically done using a narrow extraction window (red lines) to leverage the high resolution data. In contrast, when performing MRM analysis with the quadrupoles set at unit resolution, the effective width is ~1 Da as shown by the blue lines.

Reproducibility

Because the full scan data is acquired all the time, data processing decisions such as which fragment ions to extract, what extraction resolution to use (Figure 3) and whether or not summing fragment ions can improve data quality are deferred until after the acquisition is complete. This simplifies assay development, as the only parameters that need to be determined upfront are the precursors' m/z values and collision energy. Shown in Figure 4 (top) is an example for peptide VLDQLLLPYTTK from protein MRI1, where the summation of three fragment ions improved ion statistics and data quality.

The global effect of this summation of ions on improving reproducibility is illustrated in Figure 4 (bottom). The reproducibility of the raw peak areas across 3 replicate injections was measured on three fragment ion XICs (Figure 4, bottom, blue). The peak areas for the three fragment ions were summed, and the reproducibility of the summed peak area was assessed (red). The results show that better reproducibility is often achieved for low abundant proteins through the summation of fragment ions.




Figure 4. Quantitative Reproducibility on Low Abundant Proteins (50-10,000 copies/cell). Summing of ions can sometimes be used to increase the signal and therefore the quantitative reproducibility. Replicate analysis of the 43 peptides showed that the quantitative reproducibility was still high.

Multiplexing

Multiple peptides per method can be analyzed by looping together multiple MS/MS scans in the experiment, and then extracting the relevant sets of fragment ions in the data processing, as illustrated in Figure 5. In this experiment 43 different peptides from 15 proteins were targeted from the yeast proteome. These MS/MS can also be scheduled in time to further increase the numbers of peptides that can be targeted in a single run.




Figure 5. Multiplexed MRM^{HR} Workflow. 43 peptides were analyzed in a single acquisition method, the XICs of each peptide are shown here. Because full scan MS/MS data is acquired for each peptide, any number of fragment ion MRMs can be extracted for each peptide. In this case, typically 3-5 fragment ions were extracted for each peptide, corresponding to 132 XICs for quantitation.

Specificity

When targeted quantification is done on a triple quadrupole based instrument, quadrupoles are typically set at unit resolution, transmitting a 0.7 Da wide window in both Q1 and Q3. Most of the time, this resolution is sufficient to obtain high quality quantitation on peptides in complex matrices. However sometimes higher specificity is required and in these cases, performing the assays on an accurate mass system such as a QqTOF system can improve data quality. The difference with the MRM^{HR} workflow is that the TOF analyzer is used in place of 'Q3' for fragment ion detection with its inherent advantages of detecting all ions, at high resolution and high mass accuracy (Figure 1).

To illustrate, multiple fragment ions were extracted at variable extraction widths for some of the lower abundant target peptides in the yeast lysate (Figure 6). The width of the extraction window used is determined by the specificity required to get a clean quantitative peak on each peptide. In the example shown, an interfering fragment ion very close in m/z from the target y_5 fragment ion can be resolved using much narrower extraction widths. The benefit of full scan MS/MS is that selective post-acquisition extraction of the data makes it easy to establish the optimal fragment ions and XIC widths to employ to generate the highest quality quantitative data.




Figure 6. Removing Interferences or Background using High Resolution Fragment Ion XICs. The looped MS/MS for Q1 m/z 591.3 is performed across the whole LC run to target a peptide from VPS1 protein in the yeast cell lysate (A) and the total ion chromatogram (TIC) represents all the peptides with a similar nominal mass. Full scan MS/MS is present across the whole LC run and can be extracted at specific retention times (B). Zooming in on the fragment ions of interest in cases where quantitative data shows interference can reveal other fragment ions close in mass (C). The high resolution MS/MS data allows us to selectively extract the data for correct fragment ion and significantly improve the quantitative data quality (0.7 Da (pink) and 0.02 Da (blue) XIC window on the y5-ion and 0.02 Da window on interfering peak (orange) (D).

Conclusions

- The speed and sensitivity of the TripleTOF® Systems have enabled the MRM^{HR} workflow, where high resolution TOF MS/MS spectra are acquired in a looped fashion and then fragment ions are extracted post-acquisition at high resolution to generate MRM-like data. All data processing and quantitation after extraction is exactly the same as MRM data processing using MultiQuant™ Software.
- MRM^{HR} can provide better specificity and therefore higher quantitative robustness by reducing the effects of interferences.
- Assay development is simplified as full scan MS/MS is always acquired and fragment ion selection is done post-acquisition, not up-front before acquisition. Only the peptide parent ion m/z and the collision energy are required up-front, eliminating the need for time consuming method development.

References

1. Paola Picotti, Bernd Bodenmiller, Lukas N. Mueller, Bruno Domon, and Ruedi Aebersold. (2009). Full Dynamic Range Proteome Analysis of *S. cerevisiae* by Targeted Proteomics. *Cell* **138**, 795–806.
2. Ghaemmaghami S, Huh WK, Bower K, Howson R, Belle A, Dephoure N, O'shea EK, Weisman JS. (2003). Global Analysis of Protein Analysis in Yeast. *Nature* **425(6959)**, 737-41.

AB Sciex is doing business as SCIEX.

© 2017 AB Sciex. For Research Use Only. Not for use in diagnostic procedures. The trademarks mentioned herein are the property of AB Sciex Pte. Ltd. or their respective owners. AB SCIEX™ is being used under license.

Document number: RUO-MKT-02-5733-A