

Release Notes

LightSight®

Software for Metabolite Identification
Version 2.3

New Features and Changes

Includes support for the following equipment:

- QTRAP® 4500 system
- Shimadzu Nexera LC system
- Agilent 1260 LC system
- Shimadzu SIL30-ACMP autosampler

Includes support for the following operating system:

- Windows 7 (32-bit) SP1

Where to Get Help

For step-by-step animations of basic procedures, click **Help > Show Me**.

For written procedures and troubleshooting information, click **Help > Topics** or click the **How Do I?** button.

Support

- support@absciex.com
- www.absciex.com

Requirements

Processing Requirements

We recommend at least a Pentium IV 3.2-GHz system with a minimum of 1 GB of RAM. This software requires either the Windows XP SP2, Windows XP SP3, or Windows 7 (32-bit) SP1 operating system.

File Compatibility

This software supports data files created using the Analyst® software version 1.4.2 or higher.

Scan Types

This software supports the following scan types: Q1, Q3, EMS, EPI, PI, Prec, NL, and MRM.

Acquisition Requirements

- The acquisition computer must have the Analyst software version 1.4.2 or higher installed.
- The Analyst software version 1.5 cannot have been installed as an upgrade. If a previous version of the Analyst software is installed, uninstall it before installing version 1.5.
- The user cannot acquire data with the Analyst 1.4.2 Software Components for FlashQuant™ Workstation.
- For supported operating systems and computer models, see the Analyst software installation guide.

Supported Equipment

This software supports acquisition using the following equipment:

Supported Mass Spectrometers

QTRAP® 4500 System	API 5000™ System
QTRAP® 5500 System	API 4000™ System
4000 Q TRAP® System	API 3200™ System
3200 Q TRAP® System	API 3000™ System
Q TRAP® System	API 2000™ System

Supported LC Systems

Agilent 1260	Shimadzu Nexera
Tempo™	Perkin Elmer 200 series
Agilent 1290	Shimadzu Prominence series
Agilent 1200 series	Shimadzu SIL-HT series
Agilent 1100 series	
Shimadzu SIL30-ACMP autosampler	
CTC PAL autosamplers with all LC pumps supported by the Analyst software version 1.4.2, 1.5 or 1.5.1	
Waters ACQUITY UPLC system with the ACQUITY UPLC System Driver version 1.30 or 1.40	

Installation Instructions

Insert the installation DVD into the computer and follow the instructions. If the installation wizard does not open automatically, browse to the DVD and then double-click setup.exe.

If the installation files were downloaded from our Web site, browse to the folder location and then double-click setup.exe.

Note: To upgrade to version 2.3, version 2.2.1 must be installed. License activation is not required when upgrading. Earlier versions of the LightSight software must be uninstalled before installing LightSight 2.3

Known Issues and Limitations

The LightSight Software Feature Expired message

If the LightSight Software Feature Expired message is shown and the license file is valid, make sure that the network card is enabled and that the Internet Protocol (TCP/IP) setting for the network card is installed. For information about accessing network connection settings, refer to the Windows Help system.

This document is provided to customers who have purchased AB SCIEX equipment to use in the operation of such AB SCIEX equipment. This document is copyright protected and any reproduction of this document or any part of this document is strictly prohibited, except as AB SCIEX may authorize in writing.

Software that may be described in this document is furnished under a license agreement. It is against the law to copy, modify, or distribute the software on any medium, except as specifically allowed in the license agreement. Furthermore, the license agreement may prohibit the software from being disassembled, reverse engineered, or decompiled for any purpose.

Portions of this document may make reference to other manufacturers and/or their products, which may contain parts whose names are registered as trademarks and/or function as trademarks of their respective owners. Any such usage is intended only to designate those manufacturers' products as supplied by AB SCIEX for incorporation into its equipment and does not imply any right and/or license to use or permit others to use such manufacturers' and/or their product names as trademarks.

AB SCIEX makes no warranties or representations as to the fitness of this equipment for any particular purpose and assumes no responsibility or contingent liability, including indirect or consequential damages, for any use to which the purchaser may put the equipment described herein, or for any adverse circumstances arising therefrom.

For Research Use Only. Not for use in diagnostic procedures.

The trademarks mentioned herein are the property of AB Sciex Pte. Ltd. or their respective owners.

AB SCIEX™ is being used under license.

AB SCIEX

71 Four Valley Dr., Concord, Ontario, Canada. L4K 4V8.

AB SCIEX is ISO 9001 registered.

© 2012 AB SCIEX.

Printed in Canada.

