

# Certificate of Analysis

## Product Identification

Name: Standards Chemical Kit with Low/High Concentration PPGs

Part number: 4406127

Description:

BOTTLE PART NUMBER	LOT NUMBER	COMPONENT	CONCENTRATION (+/- <2%)
4405227	A3340	POS PPG 1E-4M	100 µM
4405229	A3340	POS PPG 1E-5M	10 µM
4405231	A3340	POS PPG 2E-6M	2 µM
4405233	A3340	POS PPG 2E-7M	0.2 µM
4405234	A3343	NEG PPG 3E-4M	300 µM
4405235	A3343	NEG PPG 3E-5M	30 µM
4405236	A3343	RESERPINE 1 µM	1 µM
4405237	A3353	RESERPINE 0.167 µM	0.167 µM
4405238	A3343	RESERPINE 0.0167 µM	0.0167 µM
4405247	A3345	RENIN - RESERPINE MIX (1 pmol/µL Renin + 400 fmol/µL Reserpine)	1 µM - 0.4 µM

## Lot Specific Information

Lot number: A3354

Manufacture date: 20 December 2013

Expiration date: 11 October 2014

## Quality Assurance Statement:

These standards have been manufactured using ISO9001 procedures. Concentrations are calculated gravimetrically, and balances are calibrated to NIST traceable standards in accordance with our ISO procedures. Purity from interferences is verified through mass spectrometric analysis. Solvents are minimum HPLC grade.

The information on this certificate as been reviewed and accurately reflects manufacturing and analysis data for the product lot specified.

Quality Assurance Associate: Brandon Munyon

Date: 20 December 2013

## Support

Technical Support call 1 800-952-4716 or 877-740-2129 or E-mail: [Support@sciex.com](mailto:Support@sciex.com)

This certificate may be obtained at: <http://www.sciex.com/mk/get/COFA>