

# Certificate of Analysis

## Product Identification

Standards Chemical Kit with Low/High Concentration PPGs

Part number: 4406127

## Lot Specific Information

Lot number: M011143

Manufacture date: 2020-12-18


Expiration date: 2021-09-15

Bottle Part Number	Lot Number	Component	Concentration (+/- <2%)
4405227	M010072	POS PPG 1E-4M	100 µM
4405229	M012094	POS PPG 1E-5M	10 µM
4405231	M011222	POS PPG 2E-6M	2 µM
4405233	M011113	POS PPG 2E-7M	0.2 µM
4405234	M011116	NEG PPG 3E-4M	300 µM
4405235	M009332	NEG PPG 3E-5M	30 µM
4405236	M010067	RESERPINE 1 µM	1 µM
4405237	M012103	RESERPINE 0.167 µM	0.167 µM
4405238	M012119	RESERPINE 0.0167 µM	0.0167 µM
4405247	M012072	RENIN - RESERPINE MIX (1 pmol/µL Renin + 400 fmol/µL Reserpine)	1 µM - 0.4 µM

## Quality Assurance Statement:

These standards have been manufactured using ISO9001 procedures. Concentrations are calculated gravimetrically, and balances are calibrated to NIST traceable standards in accordance with our ISO procedures. Purity from interferences is verified through mass spectrometric analysis. Solvents are minimum HPLC grade.

The information on this certificate has been reviewed and accurately reflects manufacturing and analysis data for the product lot specified.

Quality Assurance Associate: 

Date: 2020-12-18

## Support

Technical Support call 1 800-952-4716 or 877-740-2129 or E-mail: [Support@sciex.com](mailto:Support@sciex.com)

This certificate may be obtained at: <http://sciex.com/coa>