



SCIEX QTRAP 4500 MASS SPECTROMETER

***UK DECLARATION OF CONFORMITY***

**5302206**

↑  
↓

↓



**UK DECLARATION OF CONFORMITY**  
(According to BS EN ISO/IEC 17050-1)



We **AB SCIEX Pte. Ltd**  
of **Blk 33, #04-06, Marsiling Ind. Estate Road 3, Woodlands Central Industrial Estate, Singapore 739256**

declare under sole responsibility that the product as originally delivered

Equipment: **Liquid Chromatography-Mass Spectrometer/Mass Spectrometer**  
Product Name: **SCIEX QTRAP 4500**  
Product Options: **Turbo V, DuoSpray Ion Source with TurbolonSpray (TIS) or Atmospheric Pressure Chemical Ionization (APCI); and NanoSpray III Ion Source**

complies with the essential requirements of the following applicable UK legislations, and carries the marking accordingly:

- **Electrical Equipment (Safety) Regulations 2016**
- **Electromagnetic Compatibility Regulations 2016**
- **Waste Electrical and Electronic Equipment Regulations 2013**
- **The Restriction of the Use of Certain Hazardous Substances in Electrical and Electronic Equipment Regulations 2012 and applicable amendments**

The product was tested in a typical configuration and conforms with the following product standards:

<b>Category</b>	<b>Standard</b>	<b>Classification</b>
<b>Safety</b>	BS EN 61010-1:2010 +A1:2019 BS EN 61010-2-061:2021	Class I
<b>EMC</b>	BS EN 61326-1:2013	Class A

I, the undersigned, hereby declare that the equipment specified above conforms to the above Regulations and Standards.

7/13/2022

X Nichole Riek

Signed by: Riek, Nichole  
**Nichole Riek**  
**Vice President,**  
**Quality Assurance and**  
**Regulatory Affairs**