



Demonstrating reliable PFAS quantitation using EPA Method 533

Craig M. Butt¹, Holly Lee², Phil Bassignani³ and Nathan Sorelle³

¹SCIEX, USA; ²SCIEX, Canada; ³Pace Analytical

This technical note demonstrates the SCIEX novus V55 system as a powerful, next-generation mass spectrometer delivering highly accurate and precise quantitation of PFAS in drinking water using EPA Method 533. Excellent quantitative performance was achieved in laboratory fortified blanks (LFBs) with mean accuracies of 91%-128% at the 2 ng/L method reporting limit [MRL] concentration and 89-126% at 40 ng/L, and precision <20%CV across both spiking levels. A minimal system background, typically 1-2% of the MRL in the laboratory reagent blank [LRB], further enabled confident, high-quality data critical for trace-level monitoring. By leveraging the SCIEX OS Batch Automation Decision Rules, QC criteria, including carryover following the high-level calibration standard, can be automatically monitored, driving early risk mitigation, minimizing batch failures and reruns and maintaining high throughput operation.

Key benefits of PFAS analysis in drinking water using the SCIEX novus V55 system

Good retention and peak shape using the Phenomenex Luna Omega PS C18 column. Symmetrical peak shape was shown for early eluting analytes, such as PFBA and PFPeA with retention times [RTs] of 1.7 min and 2.0 min, respectively

Excellent quantitative performance in method spikes. Mean accuracies ranged from 91% to 128% at the 2 ng/L MRL level and from 89% to 126% at 40 ng/L in LFBs (n=7), with a mean precision of <20%CV across both levels

Automated QC monitoring to minimize batch failures. SCIEX OS Batch Automation Decision Rules monitored QC criteria with automated batch intervention to maintain sample throughput

Application to real-world samples. The method was applied to 7 field drinking water samples with PFBS detected greater than the MRL in one sample




Figure 1. Extracted ion chromatograms (XICs) for PFNA and PFHxS in the laboratory reagent blank (LRB) and 2 ng/L laboratory fortified blank (LFB) analyzed using EPA Method 533 on the novus V55 system.

Introduction

Per- and polyfluoroalkyl substances (PFAS) in drinking water are regulated in multiple countries and regions through either enforceable maximum contaminant levels (MCLs) or advisory guidelines. In the United States, the Environmental Protection Agency (EPA) established MCLs for PFOA, PFNA, PFOS, PFHxS, PFBS and HFPO-DA in April 2024. In addition, several U.S. states have implemented their own PFAS MCLs or guidance values, in some cases, expanding the number of regulated analytes and adopting more stringent action levels. Analysis of PFAS in drinking water is performed using liquid chromatography-tandem mass spectrometry (LC-MS/MS), commonly in accordance with EPA Method 533¹ or earlier methods such as EPA 537.1² and 537.3³. These methods employ solid-phase extraction (SPE) to concentrate and clean up water samples prior to instrumental analysis and require strict adherence to quality control criteria, including accuracy and precision requirements for laboratory water spike samples. This technical note describes the analysis of PFAS in drinking water using EPA Method 533 on the SCIEX novus V55 system. The novus V55 system is the smallest mass spectrometer in its class, 35% more compact compared to the SCIEX 5500+ system. Also, the improved energy efficiency reduces heat output and lab cooling needs by up to 40% as compared to the 5500+ system.

Methods

Sample preparation. Full method details are described in the EPA method 533 document.¹ Briefly, a 250 mL sample was preserved by adding ammonium acetate (final concentration of 1 g/L) and spiked with the suite of isotope dilution analogue standards (EPA-533APDS, Wellington Laboratories, Guelph, ON). Samples were extracted and concentrated using [Phenomenex Strata™-X-AW](#) polymeric solid-phase extraction (SPE) cartridges (500 mg/3 mL, P/N: 8B-S038-HBJ). Extracts were reconstituted in 1 mL of 80:20 (v/v) methanol/water and spiked with the isotope performance standards.

Liquid chromatography. Chromatography was performed with an ExionLC AE system using a [Phenomenex Luna C18\(2\)](#) (50 x 4.6 mm, 5 µm, P/N: 00B-4252-E0) as the delay column and a [Phenomenex Luna Omega PS C18](#) (50 x 2.1 mm, 3 µm, P/N:

00B-4758-AN) as the analytical column. The mobile phases were water (“A”) and methanol (“B”), both modified with 5mM ammonium acetate. The flow rate was 0.600 mL/min and the gradient program is presented in **Table 1** with a total runtime of 8.5 min. The column oven was set to 40°C and the injection volume was 2 µL.

Table 1: Gradient program for the analysis of PFAS in drinking water using EPA method 533 with the novus V55 system.

Time [min]	Mobile phase A [%]	Mobile phase B [%]
0.0	90	10
0.5	90	10
1.0	45	55
4.5	5	95
7.0	5	95
7.1	90	10
8.5	90	10

Mass Spectrometry. Samples were analyzed using the novus V55 system with the OptiFlow source under negative mode electrospray ionization. Data was acquired using multiple reaction monitoring mode (MRM) with one MRM transition per compound. The optimized source and gas conditions are presented in **Table 2** and the compound-specific parameters are presented in the **Appendix**.

Table 2: Source and gas conditions for the analysis of PFAS in drinking water using EPA method 533 with the novus V55 system.

Parameter	Value
Polarity	Negative
Ion source gas 1	45 psi
Ion source gas 2	65 psi
Curtain gas	45 psi
Source temperature	400°C
Ion spray voltage	-3000 V
CAD gas	10

Data analysis. The [SCIEX OS software](#) (version 5.0) was used for data acquisition and processing. Analyte raw area counts were normalized to the corresponding isotope dilution analogue (**see Appendix**). As per the EPA 533 criteria, the calibration curve was forced through the origin. The reported concentrations for the sulfonic acids were corrected for the sodium or potassium salt content. The “Decision Rules” workflow was used with the Acquisition Batch Automation feature to illustrate the automated, real-time monitoring of QC samples.

Batch decision rules

EPA Method 533 specifies that the laboratory reagent blank (LRB), analyzed immediately after the highest calibration standard, exhibits background PFAS concentrations less than one-third of the MRL. Exceedance of this threshold results in all positive results within the batch to be invalid and typically necessitates complete batch reanalysis. This causes significant time and throughput impacts for the lab. To mitigate this risk, SCIEX OS features batch acquisition “Decision Rules” that enable real-time, automated data monitoring and corrective

action [Figure 2]. First, a flagging rule (“carryover”) was created within the processing method to identify LRB results exceeding one-third of the MRL. Batch “Decision Rules” were then configured to automatically insert a double blank injection if this carryover flag failed. In this example, up to 3 consecutive double blanks were permitted; persistent failures resulted in batch stoppage. An intentionally contaminated LRB [“instrument blank [simulated]”] was evaluated to illustrate the effectiveness of this feature. These results demonstrate the flexibility of the SCIEX OS Batch Decision Rules to minimize batch failures and maintain sample throughput.




Figure 2. Acquisition batch automation decision rules workflow. An illustrative example is shown for monitoring carryover in a blank analyzed immediately after the high-level calibration standard.

Chromatographic retention and analyte separation

The combination of the Phenomenex Luna Omega PS C18 column and gradient conditions used, showed good chromatographic retention from the void volume as demonstrated by the individual PFAS compounds eluting from 1.7 min (PFBA) to 4.1 min during the 8.5 min runtime (Figure 3). Separation from the void volume, and potential co-eluting interferences is essential for maintaining good data quality. All compounds exhibited good peak symmetry using the PS C18 column and most PFAS were baseline separated. Further, clusters of branched isomers partially separated from the linear isomer peak for PFAS such as PFNA, PFDA, PFHxS and PFOS.

Sensitivity of the novus V55 system

Analysis of low-level standards showed detectable peaks at concentrations ranging from 5 to 10 pg/mL using the 2 µL injection volume (Figure 4). In general, the sensitivity of the sulfonic acids was greater than the carboxylic acids due to the slightly higher source temperature used. These results demonstrate the capability of the novus V55 system for trace level PFAS analysis of the broad analyte panel in EPA Method 533.




Figure 3. Overlaid XICs for the analysis of EPA 533 using the novus V55 system with an 8.5 min gradient. The XICs are shown for the 5 ng/mL calibration standard.




Figure 4. XICs for PFBS [5 pg/mL], PFNA [10 pg/mL] and HFPO-DA [5 pg/mL] in solvent standards using the novus V55 system with 2 µL injection volume.

Calibration standard quantitative performance: Accuracy and linear dynamic range

Calibration standards were prepared in 80:20 (v/v) methanol/water and analyzed at 7 levels from 0.25 ng/mL to 100 ng/mL (n=1). These calibration levels represent in-sample concentrations of 1 ng/L to 400 ng/L, accounting for the 250x SPE concentration factor. At the MRL level, 0.50 ng/mL, the accuracy ranged from 84.7% to 110% across the panel of 25 PFAS compounds, demonstrating good quantitative performance [Table 3]. All compounds showed good linearity throughout the calibration range with r^2 values ranging 0.987 to 1.000 [see Figure 5 for PFOA]. The only exception was NFDHA which is prone to in-source degradation and exhibited saturation in the 100 ng/mL standard. These results demonstrate the ability of the novus V55 system to accurately quantify PFAS across the 3 orders of linear dynamic range.




Figure 5. Calibration curve for PFOA from 0.25 to 100 ng/mL using the novus V55 system.

Table 3: Calibration standard quantitative performance for the analysis of EPA 533 using the novus V55 system. The data shows the accuracy of the 0.50 ng/mL MRL standard and calibration curve correlation coefficient (r^2)

Compound	0.50 ng/mL cal. std accuracy [%]	Calibration curve r^2	Compound	0.50 ng/mL cal. std accuracy [%]	Calibration curve r^2
PFBA	103	0.9992	4:2-FTS	100	0.9968
PFPeA	90.5	0.9968	6:2 FTS	110	0.9932
PFHxA	106	0.9966	8:2 FTS	95.1	0.9979
PFHpA	89.5	0.9981			
PFOA	88.7	0.9981	HFPO-DA	84.7	0.9987
PFNA	96.4	0.9976	ADONA	96.4	0.9996
PFDA	94.5	0.9978	PFMPA	102	0.9977
PFUnA	102	0.9980	PFMBA	92.7	0.9976
PFDoA	108	0.9974	NFDHA	99.8	0.9768
PFBS	92.9	0.9948	9Cl-PF3ONS	99.1	0.9990
PFPeS	102	0.9949	11Cl-PF3OUdS	89.8	0.9985
PFHxS	101	0.9972	PFEESA	95.6	0.9872
PFHpS	84.7	0.9999			
PFOS	98.6	0.9985			

Method performance in laboratory fortified blanks [LFBs]: Accuracy and precision

The method accuracy and precision were evaluated through laboratory fortified blank [LFB] spikes at 2 levels: the MRL [2 ng/L] and 40 ng/L [n=7]. In addition, laboratory reagent blanks [LRBs] were prepared and analyzed with each LFB set [n=7]. At the 2 ng/L MRL level, the mean accuracy ranged between 91%-128% [Figure 6], within the EPA 533 performance criteria of 50-150% for LFB samples spiked within 2x of MRL. Similarly, the mean accuracy ranged between 89%-126% at the 40 ng/L level, within the EPA criteria of 70-130% for LFB samples spiked at higher levels. The mean precision was <20%CV for both LFB

spiking sets except for 4:2 FTS in the 0.5 ng/L spikes [23%CV]. XICs for HFPO-DA, PFDA, 9CI-PF3ONS and PFMPA in the 2 ng/L LFB are shown in Figure 7.

The LRBs showed very low background contamination at levels less than the lowest calibration standard [0.25 ng/mL]. In general, the LRB peak areas were 1-2% of the MRL peak area.

These results demonstrate the ability of the novus V55 system to produce accurate and precise data, with very low system background, for the quantitation of PFAS in drinking water using the EPA Method 533.




Figure 6. Mean accuracy [%] and precision [%CV, error bars] for the laboratory fortified blanks [n=7] at the 2 ng/L MRL and 40 ng/L mid-level. Samples were prepared using EPA Method 533 and analyzed using the novus V55 system. Red dashed lines show the accuracy acceptance criteria.

Laboratory reagent blank (LRB)

Laboratory fortified blank (LFB) - 2 ng/L




Figure 7. Extracted ion chromatograms (XICs) for HFPO-DA, PFDA, 9CI-PF3ONS and PFMPA in the laboratory reagent blank (LRB) and 2 ng/L laboratory fortified blank (LFB) analyzed using EPA Method 533 on the novus V55 system.

Real-world drinking water sample analysis

The method was applied to 7 real-world drinking water samples. PFAS concentrations were below the MRL for all analytes except for PFBS in field sample #3 at 3.0 ng/L (**Figure 8**). These results are consistent with the generally low levels of PFAS detected in background drinking water samples. Overall, this demonstrates the capability of the novus V55 system to quantify PFAS in real-world drinking water samples using EPA Method 533.




Figure 8. Extracted ion chromatograms (XICs) for PFBS in the instrument blank field sample #3. PFBS was detected above the MRL at 3.0 ng/L.

Conclusions

The technical note demonstrated:

- Application of EPA Method 533 for PFAS in drinking water using the novus V55 system
- Good chromatographic performance using the Phenomenex Luna Omega PS C18 column; symmetrical peak shape and separation from the void-volume for early-eluting analytes (PFBA = 1.7 min, PFPeA = 2.0 min)
- Reliable quantitative accuracy and precision in laboratory fortified blank samples; with mean accuracies of 91–128% at 2 ng/L (MRL) and 89–126% at 40 ng/L (n = 7 per levels), and mean precision maintained at <20% CV across both levels
- Automated QC monitoring in the SCIEX OS software to maintain sample throughput, SCIEX OS Batch Automation Decision Rules effectively monitored QC criteria, minimizing batch failures while mitigating potential impacts
- Applicability to real-world samples; method was applied to seven field drinking water samples, with PFBS detected above the MRL in one sample (3.0 ng/L)

References

1. Rosenblum, A. and S.C. Wendelken. Method 533: Determination of per- and polyfluoroalkyl substances in drinking water by isotope dilution anion exchange solid phase extraction and liquid chromatography/tandem mass spectrometry. United States Environmental Protection Agency, Washington, November 2019. <https://www.epa.gov/sites/production/files/2019-12/documents/method-533-815b19020.pdf>
2. Shoemaker, J.A.; Grimmett, P.E. Boutin, B.K. Method 537: Determination of selected perfluorinated alkyl acids in drinking water by solid phase extraction and liquid chromatography/tandem mass spectrometry [LC/MS/MS], version 1.1. United States Environmental Protection Agency, Washington, DC, September 2009. [EPA Document#: EPA/600/R-08/092](#)
3. Shoemaker, J.A. and D.R Tettenhorst. Method 537.1: Determination of selected per- and polyfluorinated alkyl substances in drinking water by solid phase extraction and liquid chromatography/tandem mass spectrometry [LC/MS/MS], version 2.0. United States Environmental Protection Agency, Washington, DC, March 2020. [EPA Document#: EPA/600/R-20/006](#)

Appendix

Appendix: Compound specific parameters and corresponding isotope dilution analogue for the analysis of EPA Method 533 using the novus V55 system

Compound	Precursor ion [m/z]	Fragment ion [m/z]	DP [V]	CE [V]	CXP	Isotope dilution analogue
PFBA	213	169	-22	-10	-31	¹³ C ₄ -PFBA
PFPeA	263	219	-20	-8	-34	¹³ C ₅ -PFPeA
PFHxA	313	269	-25	-12	-13	¹³ C ₅ -PFHxA
PFHpA	363.1	319	-30	-14	-13	¹³ C ₄ -PFHpA
PFOA	413	369	-30	-15	-13	¹³ C ₈ -PFOA
PFNA	463	419	-45	-15	-14	¹³ C ₉ -PFNA
PFDA	512.9	469	-45	-15	-16	¹³ C ₆ -PFDA
PFUnA	563.1	519	-50	-16	-18	¹³ C ₇ -PFUnA
PFDoA	613	569	-50	-18	-20	¹³ C ₂ -PFDoA
PFBS	298.7	79.9	-80	-65	-11	¹³ C ₃ -PFBS
PFPeS	349	79.9	-90	-75	-11	¹³ C ₃ -PFHxS
PFHxS	398.7	79.9	-100	-90	-11	¹³ C ₃ -PFHxS
PFHpS	449	79.9	-120	-100	-11	¹³ C ₈ -PFOS
PFOS	498.9	79.9	-120	-110	-11	¹³ C ₈ -PFOS
4:2-FTS	327.1	307	-65	-27	-11	¹³ C ₂ -4:2-FTS
6:2 FTS	427.1	407	-80	-32	-11	¹³ C ₂ -6:2-FTS
8:2 FTS	527.1	507	-130	-37	-11	¹³ C ₂ -8:2-FTS
HFPO-DA	284.9	168.9	-20	-11	-11	¹³ C ₃ -HFPO-DA
ADONA	376.9	250.9	-22	-17	-11	¹³ C ₄ -PFHpA
PFMPA	229.1	85	-45	-10	-19	¹³ C ₄ -PFBA
PFMBA	279.1	85	-45	-10	-19	¹³ C ₅ -PFPeA
NFDHA	295.1	201	-30	-17	-15	¹³ C ₅ -PFHxA
9Cl-PF3ONS	530.9	350.9	-80	-37	-11	¹³ C ₈ -PFOS
11Cl-PF3OUdS	630.9	450.9	-120	-40	-11	¹³ C ₈ -PFOS
PFEESA	314.9	134.9	-55	-30	-13	¹³ C ₃ -PFBS

The SCIEX clinical diagnostic portfolio is For In Vitro Diagnostic Use. Rx Only. Product[s] not available in all countries. For information on availability, please contact your local sales representative or refer to <https://sciex.com/diagnostics>. All other products are For Research Use Only. Not for use in Diagnostic Procedures.

Trademarks and/or registered trademarks mentioned herein, including associated logos, are the property of AB Sciex Pte. Ltd. or their respective owners in the United States and/or certain other countries [see www.sciex.com/trademarks].

© 2026 DH Tech. Dev. Pte. Ltd. MKT-37857-A



Headquarters
250 Forest Street, Marlborough,
MA 01752 USA
Phone 508-383-7700
sciex.com

International Sales
For our office locations please call the division
headquarters or refer to our website at
sciex.com/offices